# MED DIMENSIONS

# **FELINE INTUBATION**

MD-E-F1021S



**User Manual** 

#### **Intended Use**


- 1. Use our intubation trainer to practice inserting an intubation tube or laryngoscope.Near unlimited use

#### 1. Parts Included

- 1. Head
- 2. Jaw
- 3. Top and bottom teeth
- 4. Tongue with epiglottis and pharynx
- 5. Suction cup mount

#### 2. Tools Recommended

- 1. Size 4 endotracheal tube
- 2. Laryngoscope
- 3. Lubricant



# 3. Assembly Instructions

- **1. Suction Cup Mount:** Twist the mount into the brass fitting at the back of the head.
  - **Note:** The brass fitting fits most camera tripods with a head plate screw.
- 2. Replace Teeth:
  - **a. Bottom Teeth:** Bend the tip of the jaw down and pull the teeth out. While maintaining the bend of the jaw tip, slide the replacement teeth into the jaw, release the jaw, and press on the molars to ensure proper seating.
  - **b. Top Teeth:** Reach under the nose and pull the incisors downward. Open the jaw, set the replacement teeth deep in the oral cavity, then slide the teeth forward along the hard palate, snapping the teeth into place.
- **3. Replace Jaw and Tongue:** Trim all four cable-ties and remove the jaw. Align the holes of the replacement jaw with the holes of the head. Feed the cable-ties face down through the jaw, into the head. Cinch the cable-ties until taught. Trim cable-tie tails and rotate to seat the heads of the cable-ties into the cutouts of the model.
- 4. Lubricate as you would in a normal setting. Flush with water to remove lubricant as needed.



#### **Best Practices**

To keep your models in best working order we recommend the following:

- 1. Store your model in a cool, dark, and low dust location.
- 2. For models containing or exposed to water, wipe down excess, leave all connectors open and allow components to dry completely before storing.
- 3. If applicable, apply Silicone Finishing Powder to reduce tack and dust accumulation.
- 4. If applicable, flush or wash models with water to remove lubricant and debris.
- 5. Work on a flat surface to prevent the model from slipping.

## **Replacement Components**

Component	SKU	Material	Size	Weight
Replacement Jaw/ Tongue	MD-E-F1021S-1	Plastic, Silicone	5.0x3.0x2.0 in	0.25 lbs
Replacement Teeth	MD-E-F1021S-2	Plastic	1.5x1.5x1.5 in	0.25 lbs

## **Warranty Information**

This is a summary of the Med Dimensions warranty policy provided for your convenience. For the complete and most up-to-date warranty information, please refer to the full policy available at <a href="https://www.med-dimensions.com/warranty/">https://www.med-dimensions.com/warranty/</a>.

Med Dimensions warrants its products to be free from defects in design, operation, workmanship, and materials. Warranty duration varies by type of product as described on our warranty page. This warranty does not cover damage resulting from misuse, abuse, neglect, or unauthorized repair. To initiate a warranty claim, contact Med Dimensions Customer Support at <a href="mailto:support@med-dimensions.com">support@med-dimensions.com</a> or <a href="mailto:585-510-0005">585-510-0005</a> x1. Proof of purchase is required for warranty service. This warranty supersedes all other warranties, whether expressed or implied, including warranties of merchantability or fitness for a particular purpose. Med Dimensions is not liable for consequential or incidental damages arising from product use or inability to use.

